(02/26/90)

ER PROGRAM DATA ASSESSMENT SUMMARY REPORT FORM

	ch No. <u>8909S083 - E0734</u>		
Lab	oratory <u>TMA/Eberline</u>		•
Sam	iple Numbers <u>SW070006, S</u>	SW069006, SW0	068006, SW067006, SW066006, SW031006
		Data As	ssessment Summary
y G	Gross $\alpha + \beta$ Analysis as Proportional Counters		Comments
1.	Holding Times	<u>v</u>	
2.	Initial Calibrations	<u>V</u>	See Comment 2.
3.	Continuing Calibrations	<u> </u>	
4.	Blanks		-
5.	Lab Replicates	<u>V</u>	
6.	Lab Control Samples	R	See Action Item 1.
7.	Size of Aliquot	<u>v</u>	
8.	Self Absorption-Recovery Factors	V	See Comment 3.
).	Sample Calculations	<u> </u>	See Comment 4.
0.	Overall Assessment	<u>R</u>	Data rejected.
	 V = Data had no problems. A = Data acceptable but qualified due to p R = Data rejected. X = Problems, but do not affect data. 		
Date	·	hotch ware review	ved and found to be rejected. Refer to Action Items and Comment sections
		Date I WOI O I CYIEW	rea and round to be rejected. Relet to Action fichis and Comment Sections
TOL (liscussion.		
			"REVIEWED FOR CLASSIFICATION DEVICED FOR CLASSIFICATION
			"REVIEWED FOR CLASSIFICATION REVIEWED FOR CLASSIFICATION By R. B. Hoffman By Storge H. Sotly Date 7 1/90 Date 6/27/90
			Date 11 90 Date @ 6 127/90

e0734/rk40

Action Items: 1) Lab Control Samples: The alpha results for one of two Lab Control Samples (LCSs) fell
outside 3 σ. The alpha results were flagged as estimated JA. The beta results for both LCSs fell outside 3 σ. The
beta results were flagged R (rejected). Therefore, all data were flagged R.
Comments: 1) Lab Control Samples: One out of two LCS values for the alpha analysis fell outside the 3 o
control limits; thus the data were flagged as estimated JA. Both values for the beta LCS fell outside 3 o control
limits; thus all data for gross alpha and beta were flagged as rejected (R).
2) Initial Calibrations: The gross alpha standard reference material traceable to NIST Material AMZ.340 was
diluted 11/30/83. The standard solution is at least 6 years old. The gross beta standard reference material traceable
to NIST Material 4233-B was diluted 2/9/84. The standard solution is at least 6 years old.
3) Self Absorption Factors: The gross alpha self-absorption curve for Scintillation Alpha Counters was
exceeded. E-factors, which are efficiency factors, range from 1.55 to 3.88 for the alpha curve. The E-factors for
SW069006 and SW067006 were 5.18 due to heavy dissolved solids in the sample matrix. The E-factors used for
calculation of the activities were obtained by extrapolation outside the self-absorption curve range.
4) Sample Calculations: A transcription error was found in the beta calculation for sample
SW031006. The reported value was 7.5 pCi/L, the corrected value was 5.5 pCi/L.
Note: Data Summary Tables are attached.
David WMood. 3/13/90
Reviewer Signature Date

ANALYTICAL RESULTS (pCI/L) RADIOCHEMICAL ANALYSIS

TABLE #: 8909S083 - E0734

Page 1 of 1

SITE NAME: Area 2 - Hillside	: Area 2	Hillside			
e Location					-

Sample Location		:															j									
Sample Number		SW070006	8		900690MS	006		SW068006	3006		900290MS	7006		900990MS	3006		SW031006	006		Reagent Blank	nt Bland	Reager	Reagent Blank			
Sample Date		9/26/89			9/26/89			68/92/6	Đ		68/92/6	9		9/26/89	١		9/26/89									
Matrix		Water			Mater			Water			Water			Water			Water			Water		Water				
	₽	:																								
Parameter	PCVL.	Val.	*	8	Val.	*	8	Vel.	*	8	Val.	*	8	Val.	‡	8	Val.	+/-	8	¥al.	‡	 <u>€</u>	*			
Gross Alpha	2	5	1	Ð	٨	-	Ð	9	1	æ	9	1	R	6	1	Я	6	_	B	0.1 U	0.5	25	0.3			
Gross Beta	4	12	3	R	13	3	R	27	ಚ	æ	19	2	R	9	2	R	6	2	Ŗ	c	N	20	20		\rfloor	
Total Strontium	1																									
Total Cesium	1																									
Radium 228	1																							\bot		
Tritium	400																									
Uranium 234 & 233	0.6																							_		
Uranium 235	0.6																									
Uranium 238	0.6																									
Plutonium 239 & 240	0.01																									
Americium 241 & 242	0.01																									
Radium 226	0.05																									
Gamma scan																										
Other isotopes																										

- Indicates the parameter was not detected above the Instrument Quantitation Limit
- Quantitation is approximate due to limitations identified during the quality control review
- Value is rejected due to other contractual criteria examined during the quality control review
- Value is rejected due to blank contamination identified during the quality control review
- Detection Limit in PicoCuries per Liter (pCI/L)

< 8 Data Qualifier

Valid

Acceptable with qualifications Rejected

e0734L/rk40

Rocky Flats Plant EG&G ER Program

Radiochemical Data Completeness Checklist for Radiometric and Gross α & β Analyses of Soil and Water by Gas Proportional Counters

I.	Yes Case Narrative
	Yes Abnormalities explained
	No Matrix Problems explained
	Yes Instrument problems explained
	Yes Improper collection, storage, preservation, container explained
	Yes Hold times met, explained if not met
П.	YesInitial Calibration Data Package
	Yes Detector ID
	Yes Analyst initials
	Yes Date, Time calibrated
	Yes Current Batch Date
	Yes Name, Activities, Dates of Certification of all NBS standards
	Yes Voltage settings, gain settings, or plot of voltage versus std CPMs
	Yes Plots of net std CPMs versus gain settings at voltage giving highest
	net CPM to gain ratio (crosstalk plot)
	Yes Last service or repair date for detector
III.	Yes Continuing Calibration Data Package:
	Yes Detector ID
	Yes Analyst initials
	Yes Date, Time of calibration check
	Yes Name, Activities, Dates of Certification of check standards
	Yes CPMs observed, count duration, mean counts
	No Control chart means (copy of control charts)
	Yes Background CPMs observed, results of chi square test
	No Mean of Last 10 background checks and allowable limits
	N/A Raw data from counter to verify crosstalk values
IV.	Yes Blanks Data Package
	Yes ID number of each detector the blank is counted in
	Yes Date, Times of counts
	Yes Samples and IDs in the set with the blank
	Yes Type of blank used
	Yes Detection level reported
V.	Yes Lab Replicates Data Package
	Yes Detector ID
	Yes Analyst Initials
	Yes Date, Time Analyzed
	Yes Value obtained for sample, replicates, mean values
	Yes Count Durations of samples and backgrounds
	Yes Statistical Analysis of Range, Control Limits

VI.	Yes Lab Quality Control Samples Data Package
	Yes Sample ID, Detector ID
	Yes Analyst initials
	Yes Values obtained, true value of sample
	Yes Statistical Analysis of results
	Yes Name, Activities, Certification date of QC samples
VII.	Yes Self-Absorption, Recovery Factors Data Package
	No Linear Equation for calibration curve, coefficients
	No Copy of self-absorption curve
	Yes Raw Data from counter to determine coefficients
VIII.	Yes Minimum Detectable Activity
	Yes Background measurements
	Yes Detector ID
	Yes Date, Time of count, count duration
	Yes Mean background CPM over long period
	Yes Calculated MDA for isotope of interest
IX.	Yes Size of Aliquot in Gross α & β Determination Data Package
	Yes Sample ID
	Yes Date, Time analyzed
	Yes Efficiency factor used
X.	Yes Sample Data Package
	Yes Printed report of results for sample, reruns
	Yes Computer calculations
	Yes Raw Data from counter, copies of notebook pages